



Product Service

Attestation of Conformity

No. N8A 097053 0054 Rev. 00

Holder of Certificate: Good Will Instrument Co., Ltd.No.7-1, Jhongsing Road.
Tucheng Dist.
23678 New Taipei City
TAIWAN**Product: Laboratory Equipment
(Digital Storage Oscilloscope)****Model(s): GDS-3102A, GDS-3104A, GDS-3352A,
GDS-3354A, GDS-3652A, GDS-3654A,
MSO-3102, MSO-3104, MSO-3352,
MSO-3354, MSO-3652, MSO-3654****Parameters:**
Rated input voltage: 100-240 Vac
Rated input frequency: 50-60 Hz
Rated input power: Max. 100 W
Protection class: I
IP class: IPX0
ta: 50 °C
Measurement category: CAT II
Measurement voltage: 300 V Max.**Tested according to:** EN 61010-1:2010/A1:2019
EN 61010-2-030:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.: 6111021000901**Date,** 2021-08-25
(Bill Lin)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.